

"Pre-treatment verification with film" oh no.

**Elfried Kok, Reinier de Graaf Groep, Delft
6 december 2012, Wijchen**



Content

- Situation RdGG
- Film and scanner tests
- Pre-treatment verification
 - One scan methode
- Results
- Time table
- Physics QA
- Conclusion
- Questions



Reinier de Graaf Groep

- 2 Varian Linacs 2100CD, with ExacTrac systems
- TPS: OMP and iPlan
 - IMRT
 - Stereotactic
- R&V: Mosaiq



Pre-treatment

- Stereotactic brain metastase (10-15 p.a.)
- IMRT prostate (\pm 60 p.a.) since may 2012 clinically




Film and scanner tests

- Start 2010
- Test scanner and film
 - FilmQA → FilmQA Pro
 - Epson V750 → Epson10000XL
 - EBT 2 → EBT 3
 - Scanner effects; warm-up and irregularities
 - Lateral correction
 - Homogeneity of the film
 - Filmbadges
- Test IMRT plan



Film and scanner tests

- Scanner warm-up effects
Dose 100, 200 and 300 cGy




Deviation < 0,2%
RdGG protocol 3 warm-up scans



Film and scanner tests

► Scanner effects irregularities



Film and scanner tests


- Scanner effects, irregularities

The color shift on the left side is caused by the scanner and is present in every scan. ➔ Not caused by the film.

Is this a major problem?

For analyses with the red color channel no problem (<300 cGy)

Take care with analyses; We know where the deviation is and that this is caused by the scanner.



Film and scanner tests


➤ Lateral correction

Yes there is a lateral effect

Is this a major problem?


No because we have the triple color correction.

Choose for a large scanner



Film and scanner tests

- Homogeneity of the film
 - Cut the film
 - Irradiate each piece of film with \pm 100, 200, 300 cGy
 - Scan the films
 - Measure the color values, dose values and check consistency map





3	2	1
6	5	4
9	8	7
12	11	10



Film and scanner tests


- Homogeneity of the film
 - Cut the film
 - Irradiate each piece of film with \pm 100, 200, 300 cGy
 - Scan the films
 - Measured the color values and dose values and check consistency map



Film and scanner tests


➤ Film badges

Shape of the calibration curve for different badges of film is comparable.




Pre-treatment verification

- Prostate
 - Regular calibration methode (scan time > 6 hours after irradiation)
 - One scan calibration methode (scan time \pm 1 hour after irradiation)
- Stereotactic brain metastase
 - Regular calibration method





Pre-treatment verification

- One scan method
 - Irradiate calibration patch
 - Irradiate treatment plan in QA mode Mosaiq
 - Wait about 50-60 minutes
 - Scan films
 - Analyse




Pre-treatment verification



Pre-treatment verification

➤ Analysing

- Open the box calibration file
- Place calibration regions on calibration patches (0 cGy and 209/258 cGy)
- Recalculate with triple color and Dose linear scaling
- Place fiducials
- Import 2D dose map (TPS)
- Analyse in image compare




Pre-treatment verification

➤ Results

- Regular calibration methode
- One scan calibration methode


Measurement	Point dosimetrie [% dev. In ISO to TPS]	Scan Orientation	Gamma [3%, 2 mm]	Differential Delta [3%]
IMRT Prostate 1	0.2%	Portrait	99.8%	95.0%
		Landscape	99.6%	89.0%
IMRT Prostate 2	-1.2%	Portrait	99.2%	87.2%
		Landscape	96.7%	80.3%
IMRT Prostate 3	0,6%	Portrait	99,5%	90,9%
		Landscape	99,6%	87,1%
IMRT Prostate 4	0,2%	Portrait	99,6%	96,2%
		Landscape	97,6%	85,0%

Patient ID	Datum	Standaard Kalibratie			One scan			Opmerking
		Gamma	Differential Delta	Cluster oppervlak [cm]	Gamma	Differential Delta	Cluster oppervlak [cm]	
IMRT A	4-sep	99,4%	96,3%	0,7	99,5%	96,8%	0,8	4-sep
IMRT B	4-sep	99,8%	96,8%	0,2	99,6%	93,3%	0,4	4-sep
IMRT C	11-sep	97,1%	72,3%	2,3	97,6%	77,3%	2,4	4-sep 2 Gy plan Prostate low risk
IMRT D	11-sep	99,8%	91,7%	0,1	99,8%	86,4%	0,2	4-sep
IMRT E	17-sep				99,6%	93,1%	0,4	4-sep 21 min. between cal and composite, scan after 50 min
		99,6%	94,6%	0,5	99,6%	95,7%	0,5	4-sep after 18 hour
IMRT F	17-sep				99,7%	93,6%	0,3	4-sep 12 min. between cal and composite, scan after 50 min
		99,6%	94,7%	0,5	99,8%	95,9%	0,2	4-sep after 18 hour
IMRT G	25-sep				99,3%	95,8%	0,9	4-sep scan after 1 hour
		99,4%	96,6%	0,9	99,4%	96,5%	0,9	4-sep after 16 hour
IMRT H	25-sep				99,5%	95,7%	0,6	4-sep scan after 1 hour
		99,7%	96,9%	0,4	99,7%	96,2%	0,4	4-sep after 16 hour




Pre-treatment verification

IMRT Prostate
Gamma map (3% 2 mm)




Pre-treatment verification


Stereoeccentric
Gamma map (3% 2 mm)



Dose map (3%)



Pre-treatment verificatie							
Uitgevoerd op:	12-11-2012						
Door:	Eifried						
Toestel:	Clinac						
Patient naam:	PEO test						
Patient ID:	123123						
Plan naam:	1 IMRT prostaatloge						
Energie:	X15						
Electrometer	PTW Unidos Webline, sn. 000255						
Ionisatiekamer	PTW 31010 (slangenkamer)						
Serienummer	0118						
$k_{T,P} = \frac{(T + 273,15)P_0}{(T_0 + 273,15)P}$	$T_0 = 20,0^\circ\text{C}$ $P_0 = 1013,25 \text{ hPa}$						
Temperatuur fantoom	21,4 °C						
Luchtdruk	1023,9 hPa						
$K_{T,P}$	0,994						
Calibratiefactor ionisatiekamer:	311,713 mGy/nC						
Fantoom correctiefactor Octavius:	-2,4%						
K_{tot}	319,194 mGy/nC						
Standaard veld (MLC 10x10 cm)	Koehc CAL4 128MU						
Dosis TPS	100,44 cGy						
Gemeten output [pC]	3157 3157						
Berekende dosis	100,20 cGy						
Correctie dosimetrie	-0,24%						
Film dosimetrie	JA, one scan methode	Kal. strook 6					
Badge nummer:	A04041202	Doos nr.: 21					
Verschuiving ISO:	NEE						
Resultaten puntdosimetrie							
Bundel	ME planning	Dosis uit Iplan TPS [cGy]	Metingen [pC]	Gemeten [cGy]	Afwijking t.o.v. berekend	Gemeten corr. dosim. [cGy]	Afwijking t.o.v. berekend
Bundel 1	144	37,70	1174	37,26	-1,2%	37,35	-0,9%
Bundel 2	123	41,71	1267	40,21	-3,6%	40,31	-3,4%
Bundel 3	129	38,15	1208	38,34	0,5%	38,43	0,7%
Bundel 4	135	39,11	1240	39,36	0,8%	39,45	0,9%
Bundel 5	129	41,57	1271	40,34	-3,0%	40,44	-2,7%
Totaal	660	198,24	6160	195,51	-1,4%	195,98	-1,1%
Criteria voldaan							JA
Resultaten Filmdosimetrie							
Film ID	Kleurkaal	Gamma [3%, 2 mm]	Dosis match	Cluster opp. Gamma > 3%, 2 mm [cm²]	Criteria voldaan		
4443402	Rood	99,6%	95,8%	0,5	JA		
Resultaten RadCalc							
	Plan	Maximale afwijking bundels	Criteria voldaan				
Afwijking RadCalc vs TPS	-0,1%	-2,3%	JA				
Nader onderzoek plan door Klinisch Fysica RT	NEE						
Plan akkoord:	Eifried Kok						
Datum:	12-11-2012						
Opmerking:							



Report Physics, Radiation therapy RdGG



Pre-treatment verification

➤ Time table


Controle	Regular calibration	One scan calibration
preparation	30 minutes	30 minutes
Linac time	30-45 minutes	20 minutes
FilmQA Pro	20 minutes	10 minutes
Administration	10 minutes	10 minutes
Total	90-105 minutes	70 minutes



Physics QA

➤ Picket fence


- Slit 2 mm with MLC
- 11 slit positions with distance from 2 cm
- 100 MU/ slit position
- Film in collimator adapter (67,5 cm)



Physics QA

➤ Picket fence




- Slit 2 mm with MLC
- 11 slit positions with distance of 2 cm
- 100 MU / slit position
- Film in collimator adapter (67,5 cm)




Physics QA

➤ Picket fence

- Slit 2 mm with MLC
- 11 slit positions with distance of 2 cm
- 100 MU / slit position
- Film in collimator adapter (67,5 cm)
- Scan film
- Analyse
 - Set the MLC design
 - Set Δ average 2,0 mm
 - Leaf opening 2,3 mm
 - Bias 0,3 mm
- report





➤ Picket fence





Physics QA

➤ Starshot

- Slit 1 mm with MLC in centre
- 120 MU / slit position
- Collimator, gantry and table
- Film in Octavius adapter

Veld	Collimator hoek
1	198°
2	234°
3	270°
4	306°
5	342°
6	0°
7	36°
8	72°
9	108°
10	144°



Physics QA

- Starshot (spaken film)
 - Slit 1 mm with MLC in centre
 - 120 MU / slit position
 - Collimator, gantry and table
 - Film in Octavius adapter
 - Set fiducials
 - Analyse
 - tool
 - Adjust alignment
 - report



Physics QA

➤ Starshot










Image
Source
Size
Resolution
Origin
Analysis color

Start shot #1
EPSON Expression 10000XL
(W= 586, H= 719) pix = (W= 206.7, H= 253.6) mm
72 dpi
RDGG\koke 6-11-2012 14:48:21
Red (beam darker)



Charts 'Dose film' star shot analysis



star shot statistics (red)

isocenter (centroid, 35 o: X=103.5, Y=126.2) mm (X=293, Y=358) pix	2 pix
dist from machine isocer 0.7 mm	2 pix
center of outer circle, 35 (X=103.4, Y=126.1) mm (X=293, Y=357) pix	
radius outer circle, 35 of 0.4 mm	1 pix
offset circle center, 35 of 0.7 mm	2 pix
ave dist to circle center, 0.2 mm	1 pix
angles 1+	144.2 °
angles 3+	180.1 °
angles 5+	216.0 °
angles 7+	252.0 °
angles 9+	288.1 °



Wish list

- Correction for scanner irregularities
- Winston Lutz analyses
- Extensive manual




Conclusion

**The question:
"Pre-treatment verification
with film"**



Questions?



Thanks

Special thanks:
Andre Micke

for all the answers to our questions and
for the tools built in the software, to make
our (filmdosimetry) work easier. 



thanks for the invitation

